

Forwarded document :

Certificate of Analysis

| | |
|------------------------|---|
| Product | Compact Dry ECO (EC only) for <i>E. coli</i> |
| ID-No. | 1 700 001 /1 700 002/1 700 003 1 600 001 |
| Lot Number | 019906 |
| New Expiry Date | 30.11.2020 |

| Test | Specifications | Results* | |
|-------------------------------|--|-----------------------|-------------|
| Appearance | Sheet of light yellow colour | Corresponds | |
| Loss on drying | < 10,0 % | Corresponds | |
| Aseptic test | No growth of any bacteria (30°C/5 days, Visual check) | Corresponds | |
| pH value | 6.8 – 7.2 | Corresponds | |
| Colony growth | | | |
| <i>Escherichia coli</i> | ATCC8739 | Bluish green colonies | Corresponds |
| <i>Escherichia coli</i> | ATCC25922 | Bluish green colonies | Corresponds |
| <i>Pseudomonas aeruginosa</i> | ATCC9027 | White colonies | Corresponds |
| <i>Klebsiella oxytoca</i> | ATCC13182 | White colonies | Corresponds |

This lot meets the specifications.

Date: 21.09.2019

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.